



Dr. Somnath Bhattacharyya
Professor in Genetics and Plant Breeding
Officer in Charge, Crop Research Unit
Research Directorate

e.mail: somnathbhat@yahoo.com,
bhattacharya.somnath@bckv.edu.in
Mohanpur, Nadia, WB 741252, India

Bidhan Chandra Krishi Viswavidyalaya,

e-Tender Notification

Ref. No. BCKV/CRU/2526/P I (ICAR)/Ten 3

Date. 25.06.25

Officer in Charge, Crop Research Unit, and PI, Adhoc-funded project is inviting price-quotation from the competent and *bonafide* vendors/ dealers having registration of GST for the supply of the following to the University within **seven days** from the date of publication of this notice in the News Paper as per specifications appended below (Table 1). **Vendors are requested to tender technical and financial bids separately, and the quoted price must include GST and other taxes if any and transportation & installation costs up to the laboratory in Mohanpur.** Price quotations will not be considered for financial comparison if they do not match the detailed specifications and warranty period described in Table 1 below. The payment will be processed only after satisfying the supplied materials as per the desired quality and quantity in the tender document and the order issued.

Table 1: Ref. No. BCKV/CRU/2526/P I (ICAR)/Ten 3

Date. 25.06.25.

S.No	Name of the Item	Qty	Detailed Specification	Cost
1	Dryer Oven cum incubator shaker with illumination system	1	<p>Dryer Oven cum incubator shaker with illumination system: Inner Size must be 18cubic ft or above;</p> <p>The unit features a triple-walled construction, where the inner chamber is made of heavy-gauge stainless steel (316 quality) with PUF insulation between the walls. The outer chamber must be made of SS304 grade finish. Temperature range from 05°C to 70°C accuracy $\pm 1^\circ\text{C}$ with a hermetically Danforst make sealed CFC free refrigerant compressor. Shaker tray (Double tray) fitted on the bottom with 30" x 30" stainless steel lotus clamps where 30x500ml Conical flasks must be accommodated.</p> <p>Variable shaking speed 10 to 400 rpm with digital speed regulator. Heavy duty DC motor capable of long duration operation (continuous 5 days minimum). Solid state controller with PT-100 sensor for controlling the Temperature and illumination system (up to 20000 Lux). Microprocessor controller with 4 LCDs of shaking speed and temperature, nine program memories, power failure and resumption recorded with data and time, Safety cut-off alarms for high and low temperature, as well as illumination.</p> <p>On-site warranty for the whole equipment, sensors and accessories: 03 years. The manufacturer must be CE & ISO 9001-2015 certified, and Technical documents must be attached separately along with the bid. If the bidder is a manufacturer, then the manufacturing licence must be enclosed. After full compliance with the desired technical specifications, the financial bid will be opened.</p>	

Somnath Bhattacharyya
25/6/25

As per decision of the 77th meeting of LTC held on 25/06/2025 tender notice is allowed to publish as per rule.
Subhrakul Singh 25/6/25

(S Bhattacharyya)