



BIDHAN CHANDRA KRISHI VISWAVIDYALAYA

Directorate of Research, Kalyani, Nadia, West Bengal

Ref. No.: BCKV/PP/SD/Instrument tender - 01/25-26

Date : 02.09.2025

NOTICE INVITING TENDER

- Principal Investigator of the ad-hoc Project “Evaluation of Bio-efficacy and Phytotoxicity of IIF – 123 against late blight disease of tomato” (Code: 30768) is inviting tenders from competent and *bonafide* manufacturers/vendors/parties/distributors/dealers/agents having registration of GST for supply of the following item to the above mentioned project, being implemented under the Directorate of Research, BCKV, Mohanpur and located at Plant Pathology Department, F/Ag., BCKV, Mohanpur, Nadia, within 14 days (up to 4.00 pm) from the date of publication at BCKV website, as per specifications appended at annexure-I. Sealed quotations in closed envelope and writing the tender reference number on it, to be submitted to Dr. Subrata Dutta, Professor & Principal Investigator, ad-hoc project entitled “Evaluation of Bio-efficacy and Phytotoxicity of IIF – 123 against late blight disease of tomato” (Code: 30768), Department of Plant Pathology, F/AG., BCKV, Mohanpur-741252, Nadia, West Bengal.

Sl. No.	Item	Specification
1.	<u>UV-Transilluminator</u>	Attached in Annexure -A
2.	<u>Vortex</u>	Attached in Annexure -A
3.	<u>Electrophoresis Gel Unit</u>	Attached in Annexure -A

Terms and Conditions:

- * Quoted price should be included of the costs of GST, accessories and other charges (delivery, installation, demonstration, etc.), if any. It should be in INR (*No provision of foreign currency transaction*).
- * The bids should be prepared and submitted with validity period of minimum 31.03.2026.
- * The bidders must stipulate the delivery period of the same within 15 days at Dr. Subrata Dutta, Department of Plant Pathology, F/Ag., BCKV, Mohanpur, Nadia, West Bengal, Pin - 741252 from date of issuing supply orders and to specified place of delivery at office of working days. (time - 11.00 am to 05.00 pm)
- * Any deviation from required specifications and standards will be treated as cancellation of bids at any stage and also cancellation of supply order, even after delivery, without further communication/clarification and no claim will be entertained thereof.
- * EMD is applicable (1%) of total bid amount in the form of Bank Draft in favour of 'Bidhan Chandra Krishi Viswavidyalaya' payable at SBI, Kalyani Branch (IFSC: SBIN0001082). Photocopy copy of the demand draft to be submitted as supporting document during submission of tender and the original copy of above mentioned demand draft to be submitted to the Principal Investigator prior to receive supply order. Document and certificate must be attached against the claim of EMD exemption, if any. Proper documents to be submitted for EMD exemption claim, if any.

- * Challan and bill should be submitted in triplicate and bank account details (Name, A/C No., IFSC & MIRC codes) of the beneficiary to be mentioned in bill.
- * Bill will be processed after delivery, installation and successful demonstration.
- * The vendor's experience on supplying to Government / Semi Government organization is required.
- * Supporting documents (as may be applicable):
 - (a) * Bid papers should accompany authorization certificate from original manufacturer, GST registration, latest IT return, PAN and other statutory documents, if any.
 - (b) * Scanned copy (self-attested) of the original supporting document in favour of the specification - claim for each item must have to be attached separately.
 - (c) * User list along with certificate from reputed users also need to be attached.
- * The Viswavidyalaya reserves the right to accept or reject any tender without showing reason.

Thanking you,


(DR. Subrata Dutta)

As per decision of the 81st meeting of LTC held on 22.09.2025, tender notice is allowed to publish as per int.

Subrata Dutta
22/9/25

1) General Specifications for UV Transilluminator

UV source : 312 nm Wave Length

Gel size : 200 x 200 mm

Lid : UV Protection Lid / Shield

- Warranty : 24 months

2) General Specifications for Vortex

i. Maximun speed : 3000 rpm

ii. Efficient vortexing : upto 15 ml tube

iii. Shaking : Orbital

- Warranty : 12 months

3) General Specifications for Mini horizontal gel running system

Gel Tray Size : 7 x 10 cm, to be supplied with buffer tank with lid, electrodes and cables, 6 and 10 well combs , Gel caster for tape free casting.

- Warranty : 12 months

Signature
04/07/15

Principal Investigator