



Bidhan Chandra Krishi Viswavidyalaya
Directorate of Research, Kalyani - 741 235, Nadia, West Bengal
AICRP on Soil Test Crop Response Correlation
☎ - (033) 9433777855 (M): 2582-8354 (R): Fax: (033) 2582-8450
E-mail: nihar_bckv@rediffmail.com

Dr. Niharendu Saha
Associate Professor & Officer in-charge

Ref No. STCR/1041/17

Date: 10.01.2017

NOTICE

QUOTATION FOR THE PURCHASE OF THE LISTED ITEM

The following item with specifications listed below are required for the research Project "AICRP on Soil Test Crop Response (Code No.924)", running at the Directorate of Research, BCKV, Kalyani, Nadia. Sealed quotations are invited from the competent and bonafied parties/distributors/dealers/agents/ manufacturers having registration of VAT for the best offer price to supply the following item to the Laboratory of the undersigned on or before 31.01.2017 by 5 pm in a sealed envelope. In case of Manufacturer, a Manufacturing/ Proprietary certificate is essential or a certificate of the authorized agent from the manufacturer is required. Vendors are requested to tender the item in prescribed format only. The Purchase Committee reserves the right to reject any or all quotations without assigning any reasons. Delivery and Installation charges are free of cost. The Rate must be given inclusive of VAT and other taxes if applicable. The validity of the quotation will be 31st March, 2017. Please note that 10 per cent of the total value will be kept in hold as security deposit for 1 year from successful installation at site.

Niharendu Saha
Dr. Niharendu Saha 10/01/17
Officer-In-Charge
AICRP on STCR

TENDER NOTICE

Sealed tenders are invited from the interested vendors for the following item:

	Technical Specifications
Research Centrifuge (1 No.) with Rotor Head (1 No.) with Tube Size 100 ml	Max. speed 17300 RPM
	Max. RCF 27440
	Max. Tube size 100 ml. (angular rotor head)
	Tube should be plastic
	Max. Capacity 400 ml.
	Stepless speed regulator with zero start interlock
	Digital speed indicator
	Dynamic Break
	0-99 minutes digital countdown timer
	Imbalance detector with cut off
	safety lead interlock to prevent cover opening during centrifuging

Interested parties are requested to quote their rates on or before 3-00 pm of 31.01.2017 in sealed cover addressed to: **Dr.Niharendu Saha, Officer-In-Charge, AICRP on STCR, Directorate of Research, BCKV, Kalyani, Nadia, W.Bengal-741235.**

The undersigned is not bound to accept the lowest or any other quotation and reserves the right to reject any or all the quotations received in this connection without assigning any reason what so ever. The quotation reference number and subject must be mentioned on the top of envelope.

Sd /

Officer-In-Charge
AICRP on STCR

Date: 10.01.2017

Ref. No:

Copy forwarded(with the request for wide circulation/Display in the notice board) to :

1. Registrar, BCKV, Mohanpur, Nadia.
2. The In charge, Comptrollers Department, BCKV, Mohanpur, Nadia.
3. Notice Board: Dean, F/ Agri, F/ Hort, F/ Agri. Engg..
4. Notice Board: Administrative Building, BCKV, Mohanpur, Nadia.
5. Notice Board: Common Service Section, BCKV, Mohanpur, Nadia.
6. Notice Board: Office of the Director of Research, Research Complex, Kalyani, Nadia.
7. Dr. Chandan Karak, BCKV, Mohanpur, Nadia. He is requested to upload the Matter to the BCKV website as early as possible.
8. File Copy

Niharendu Saha
(Dr. Niharendu Saha) 10/01/17

Officer-in-charge
AICRP on S.T.C.R.
Directorate of Research
Bidhan Chandra Krishi Viswavidyalaya
Kalyani-741235, Nadia, West Bengal