



Bidhan Chandra Krishi Viswavidyalaya
Directorate of Research, Kalyani - 741 235, Nadia, West Bengal

AICRP on Soil Test Crop Response Correlation

☎ - (033) 9433777855 (M): 2582-8354 (R): Fax: (033) 2582-8460

E-mail: nihar_bckv@rediffmail.com

Dr. Niharendu Saha

Associate Professor & Officer in-charge
PI, RIDF project on "Farmers' Service
Centre for Soil Testing and fertilizer.."

Date- 29.01.18

Ref. No. STCR/RIDF/ FSC/ II Tender/03/IV

Ref. e-Tender ID: 2018_BCKV_151769_1 Dated: 30.1.2018

Notice Inviting E-Tender (2nd Time)

The Principal Investigator, RIDF project "Farmers' Service Center for Soil Testing and Fertilizer Recommendation-referral laboratory", AICRP on Soil Test Crop Response Correlation, Directorate of Research, Bidhan Chandra Krishi Viswavidyalaya, Kalyani is inviting e-tender from competent and bonafide manufacturers/ authorized distributors/ dealers/ agents for supply of following scientific equipment to University as per specifications appended against each item.

(i) Preparation of Bids: - The tenders should be submitted under two-bid system (i.e. Technical bid and Financial bid) with validity for a period of six month or more.

(ii) Submission of e-tender: - The tender shall be submitted strictly **within 19th February, 2018 at 5 P.M; Quotation with more than one equipment in a single sheet will not be considered for comparison.** Technical Quotations may also be submitted **separately for each item** by sealed envelope mentioning 'name of the item' and 'item reference number' on top of the envelope to the below-mentioned addresses.

For **on line submission** of Tender please visit [https:// wbtenders.gov.in](https://wbtenders.gov.in)

(iii) Opening of the tender: - The technical bid will be opened by the Central Purchase Committee, BCKV duly consigned for this purpose. After scrutinizing, the financial offer/bid of technically qualified vendors will be opened.

(iv) Prices: - The prices of all the equipment and items, including imported ones, should be quoted in net per unit (including all taxes and duties, etc.). However, University will provide valid DSIR and authorization certificate to the clearing agent, if required. Quotation must include essential all accessories (as mentioned for each equipment) from branded company, in **Indian Rupees** only, even if the main equipment is quoted in foreign currency. Without essential accessories tender will be considered as incomplete. Quoted rates must be FOR DESTINATION (including packing, insurance and delivery charges up to the **Referral laboratory at BCKV, Mohanpur**) with satisfactory installation and demonstration.

(v) EMD: - Vendors are required to pay the requisite amount 4000.00 (four thousand only) as EMD in draft as specified **against each equipment/goods individually (item wise separately)**. Scan copy of the requisite draft must be uploaded as the supporting document for each equipment during submission of e-tender. **Without EMD quotations will not be considered for technical or financial comparison.** Draft must be in favour of '**Bidhan Chandra Krishi Viswavidyalaya**' payable at Kalyani (IFSC: SBIN0001082).

(vi) **Supporting documents:-**

(a) Bid papers should accompany authorization certificate from original manufacturer, trade license, GST registration, etc;

(b) Scan copy (self attested) of the original supporting document in favour of the specification-claim for each item must have to be uploaded separately;

(c) User list along with certificate from reputed users also need to be uploaded for each item;

(d) Scan copy of supporting document of assured after sales services in Eastern India and availability of spare parts need to be uploaded.

Price bid of the vendors will be compared only if technical specificity as appended against each item is fulfilled. The Viswavidyalaya authority reserves the right to accept or reject any tender without showing any reason.

Other conditions

1. The each item should be quoted with **4 years comprehensive AMC after expiry of standard warranty of 1 year (Mandatory for each item).**
2. Must have post sales local service back up facilities so that problem could be attended within 48 hours (documents to be attached)

Dr. Niharendu Saha, Associate Professor,

Principal Investigator, RIDF project

Farmers' Service Centre For Soil Testing And Fertilizer

Recommndations (Referral Laboratory)

&

Officer-in-Charge,

AICRP on Soil Test Crop Response Correlation

Directorate of Research,

Bidhan Chandra Krishi Viswavidyalaya, Kalyani-741235

Email: nihar_bckv@rediffmail.com Mobile: 8013472996, 09433777855,

List of Scientific Equipment with specifications:

Sl. No.	Item Ref. No.	Instruments
1	FSC/IITender/Instrument/Item.11	Desiccators with vacuum pump
2	FSC/II Tender/Instrument/Item.13	Fume Hood

Sl. No.	Item Ref. No.	Instruments	Quantity	Details Specification
1	FSC/IITender/Instrument/Item.11	Desiccators with vacuum pump	1	Desiccators (vacuum), Poly carbonate Top & Chemical Resistant, Polypropylene bottom, Silicon 'O' ring to make the the desiccators airtight; Fitted with stop cork made by polypropelene, Oil free
2	FSC/II Tender/Instrument/Item.13	Fume Hood	1	For digestion of different types of sample for analysis, including water, wastewater, food products, plant tissue, feed stuffs, oil, coal, organic & mineral samples for analysis of Kjeldahl nitrogen, protein, phosphorus, calcium, potassium, lead, cadmium etc in solid and liquid samples. Fume Chamber must withstand tri-acid fume constantly System . This should be alternate to conventional digestion eliminating the need for a mercury or metal catalyst. -The apparatus set should have a heater assembly heat shield, fractionating column, capillary funnel, digestion flasks and a water aspirator and Should have a thermostatically controlled 250 watt heating element and should provide even heating throughout the digestion process and have variable temperature control from 100 to 480 deg C with LED digital display for temperature