



Dr. Somnath Bhattacharyya
 Professor in Genetics and Plant Breeding
 Officer in Charge, Crop Research Unit
 Research Directorate
 Bidhan Chandra Krishi Viswavidyalaya,

e.mail: somnathbhat@yahoo.com,
 bhattacharyya.somnath@bckv.edu.in
 Mohanpur, Nadia, WB 741252, India

e-Tender Notification

Ref. No. BCKV/CRU/2425/P I (ICAR)/Ten 8

Date. 20.12.24

Officer in Charge, Crop Research Unit, and PI, Adhoc-funded project is inviting price-quotation from the competent and *bonafide* vendors/ dealers having registration of GST for the supply of the following to the University within **seven days** from the date of publication of this notice as per specifications appended below (Table 1). **Vendors are requested to tender technical and financial bids separately, and the quoted price must include GST and other taxes if any and transportation & installation costs up to the laboratory in Mohanpur.** Price quotations will not be considered for financial comparison if they find a mismatch with detailed specifications and warranty period described in Table 1 below. The payment will be processed only after satisfying the supplied materials as per the desired quality and quantity in the tender document and the order issued. *Vendors are also required to be attached the necessary trade license*

Table 1: Ref. No. BCKV/CRU/2425/P I (ICAR)/Ten 8

Date. 20.12.24

S.No	Name of the Item	Qty	Detailed Specification	Cost
1	Electro blot apparatus: Wet Transfer Apparatus:		Electro blot apparatus: Wet Transfer Apparatus: Should transfer two 10 x 7.5 cm gels simultaneously; should have high field strength for rapid 1 hr transfers; should be able to run overnight at low voltage; Electrode-wires should be placed at least 4 cm apart for strong electrical fields and efficient protein transfer Colour-coded cassettes and electrodes for proper orientation of the gel during transfer; Must have Ice cooling unit, completely contained within the apparatus, for absorbing heat generated during rapid transfers; Volume of buffer should not exceed 450ml. Powerpack basic: Output specifications 250 V, 3.0 A, 300W Output range (programmable) 10-300 V, fully adjustable in 1 V increments 4-400 mA, fully adjustable in 1 mA increments 75 W, (maximum) Type of output Constant voltage or current, with automatic crossover; Output terminals: 4 pair recessed banana jacks in parallel Timer: Up to 99 hr, 59 mm should have pause/resume function Operating condition: 0-40°C, 0-95% humidity in the absence of condensation	

As per decision of the 64th meeting of LTC held on 20.12.2024 tender notice is allowed to publish as per rule

Somnath Bhattacharyya
 20/12/24
 (S Bhattacharyya)

Somy
 20.12.24

e-tender no. 2025- BCKV- 812716-1
 of 7/02/2025

Librarian

Phy 11/2/25

21/2/25