



# BIDHAN CHANDRA KRISHI VISWAVIDYALAYA

Directorate of Research, Kalyani, Nadia, West Bengal

Ref. No. : SMLL/Purchase/2025-26/02

Date: 09-01-2026


## NOTICE INVITING TENDER

Principal Investigator of the ad-hoc Project “Evaluation of Bio-efficacy and Phytotoxicity of new fungicide, KKK-7 against Downy mildew of Grapes” (code-30900) is inviting tenders from competent and *bonafide* manufacturers/vendors/parties/distributors/dealers/agents having registration of GST, PAN Card, MSME (if any) etc. for supply of the following item to the above mentioned project, being implemented under the Directorate of Research, BCKV, Mohanpur and located at AICRP on MAP & Betelvine, Directorate of Research, BCKV, Mohanpur, Nadia within 10 days (up to 4.00 pm) from the date of News Paper and at BCKV website (which ever is later) as per specifications appended below. Sealed quotations in closed envelope and writing the tender reference number on it, to be submitted to Dr. Goutam Mondal, Professor & Principal Investigator, ad-hoc project entitled “Evaluation of Bio-efficacy and Phytotoxicity of new fungicide, KKK-7 against Downy mildew of Grapes” (code-30900), AICRP on MAP & Betelvine, Directorate of Research, BCKV, Mohanpur, Nadia-741252, WB.

Sl. No.	Name of the Item/Equipment	Detail Specification
1.	BOD Incubator with accessories	<p><b>Specification:</b> 10cft capacity; triple walled of SS 304 inner chambers material and heavy gauge – powder coated GI sheet outer chamber; Interior illumination on door with a cyclic timer to control cyclic illumination conditions, 5°C to 60 °C ambient temperature, <math>\pm 0.5^{\circ}\text{C}</math> temperature accuracy and <math>\pm 2.0^{\circ}\text{C}</math> temperature uniformity; 55 % to 90% <math>\pm 5</math> % RH with Digital Humidity Controller/ Indicator Humidity Control arrangement; microprocessor based digital temperature indicator-cum-controller with LED OR LCD display for set value (SV) and process value (PV); forced air circulation at the triple-walled back by the durable coaxial blower/fan Air Circulation; 3-4 removable, sturdy, elegant-precisely fabricated, minimum resistance to air circulation shelves of SS wire mesh; nearly silent operation with ultra-low vibration; digital volt meter fitted on panel for Voltage Indicator; 220V AC, 50Hz Electric Supply; auto power break off, over temperature protection &amp; safety alarm system.</p> <p><b>Warranty:</b> Five years full comprehensive warranty from the date of installation.</p>

### **Terms and Conditions:**

- \* Quoted price should be included of the costs of GST, accessories and other charges (delivery, installation, demonstration, etc.), if any.
- \* Bill will be processed after delivery, installation and successful demonstration.
- \* The Viswavidyalaya reserves the right to accept or reject any tender without showing reason.



09/01/2026

(DR. Goutam Mondal)

As per decision of the 87th  
meeting of the LTC held on..09/01/2026  
tender notice is allowed to publish  
as per rule

Subhankar K. Mondal  
09/01/2026

Prof. Goutam Mondal  
Principal Investigator  
Evaluation... fungicide, KKK-7...Grapes  
Directorate of Research, BCKV  
Mohanpur-741252, Nadia, W.B.